



## CE DECLARATION OF CONFORMITY

**MANUFACTURER'S NAME:** Entegris Inc.

**MANUFACTURER'S ADDRESS:** 101 Peavey Road, Chaska MN 55318

**EQUIPMENT DESCRIPTION:** Pressure Transducers

**EQUIPMENT MODEL DESIGNATION:** 4100, 4210

**APPLICABLE DIRECTIVES:** EMC Directive: 2014/30/EU  
ATEX Directive: 2014/34/EU  
ROHS Directive: 2011/65/EU

**APPLICABLE STANDARDS:**

Emission Requirements:

EN 61326-1: 2013, Class B

EN 55011: 2009+A1: 2010, Class B

Immunity Requirements:

EN 61326-1: 2013

EN 61326-2-3: 2013

EN 61000-4-2: 2009

EN 61000-4-3: 2003 + A1: 2008 + A2: 2010

EN 61000-4-4: 2004 + AC: 2006 + A2: 2010

EN 61000-4-6: 2009

EN 61000-4-8: 2010

ATEX Requirements:

EN 60079-0: 2018

EN 60079-11: 2012

ROHS Requirements:

EN IEC 63000:2018

**EMC TEST REPORT NUMBERS:** NC1407453.1 & NC1407453.2

**ATEX TEST REPORT NUMBER:** 104274744DAL-002

**RESPONSIBLE PERSON:** Tom Shurson

**TITLE:** Vice President, Global Quality

**THIS IS TO CERTIFY THAT THE REFERENCED PRODUCT HAS BEEN DESIGNED AND MANUFACTURED TO MEET THE DIRECTIVES AND THE STANDARDS LISTED ABOVE.**

Tom Shurson

Title: Vice President, Global Quality

**In accordance with EC NOTICE TO STAKEHOLDERS  
WITHDRAWAL OF THE UNITED KINGDOM AND EU RULES IN  
THE FIELD OF INDUSTRIAL PRODUCTS dated 13 March 2020.**

This issued certificate - Certificate No: ITS20ATEX205536X

and supporting Technical Construction File underwent a legal transfer of new ownership by signed agreement between the named applicant on this certificate and the 3<sup>rd</sup> party bodies involved in the transfer from NB0359 to NB2575 on 31 December 2020

**Name:** Fabrizio Massei

**Position:** ATEX Certification Officer

**Signature:**

**Date:** 31 December 2020

12/17/2020

Date